



EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer:

Micromeritics Instrument Corporation
4356 Communications Drive
Norcross, GA 30093, USA

Hereby declares that the product:

MasterTech 052 Automatic Sample Feeder

is in conformity with the following **EU harmonization legislation**:

2014/35/EU - LVD Directive
2014/30/EU - EMC Directive
2011/65/EU - RoHS Directive

and that the equipment is in conformity with the following harmonized and other appropriate standards;

2014/35/EU (LVD)

EN 61010-1:2010/A1:2019 - *Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements.*

EN IEC 61010-2-081:2020 - *Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes*

2014/30/EU (EMC)

EN 61326-1:2013 - *Electrical equipment for measurement, control and laboratory use — EMC requirements — Part 1: General requirements*

EN 61000-3-2:2014 - *Part 3-2: Limits — Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)*

EN 61000-3-3:2013 - *Part 3-3: Limits — Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection*

2011/65/EU (RoHS)

EN 63000:2018 - *Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances*

Name: **Samuel Varner, Ph.D.**

Title: **Director, R&D**

Signature: 

Date of issue: **2026-05-26**

Location: **Norcross, GA USA**