



## EU DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer:

**Micromeritics Instrument Corporation**  
**4356 Communications Drive**  
**Norcross, GA 30093, USA**

Hereby declares that the product:

**TriStar II Plus Gas Adsorption Analyzer**  
**Model 3030**

is in conformity with the following **EU harmonization legislation**:

**2014/35/EU - LVD Directive**  
**2014/30/EU - EMC Directive**  
**2011/65/EU - RoHS Directive**

and that the equipment is in conformity with the following harmonized and other appropriate standards;

**2014/35/EU (LVD)**

**EN 61010-1:2010+A1:2019** - *Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements.*

**2014/30/EU (EMC)**

**EN 61326-1:2013** - *Electrical equipment for measurement, control and laboratory use — EMC requirements — Part 1: General requirements*

**EN 61000-3-2:2014** - *Part 3-2: Limits — Limits for harmonic current emissions (equipment input current  $\leq 16$  A per phase)*

**EN 61000-3-3:2013** - *Part 3-3: Limits — Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current  $\leq 16$  A per phase and not subject to conditional connection*

**2011/65/EU (RoHS)**

**EN 63000:2018** - *Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances*

**Name:** Samuel Varner, Ph.D.

**Title:** Director, R&D

**Signature:** 

**Date of issue:** 2026-05-26

**Location:** Norcross, GA USA