

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product

Semi-Automated Laboratory Analysis Equipment

Name and address of the applicant

Micromeritics Instrument Corporation  
4356 Communications Drive  
Norcross, GA, 30093  
United States of America

Name and address of the manufacturer

Micromeritics Instrument Corporation  
4356 Communications Drive  
Norcross, GA, 30093  
United States of America

Name and address of the factory

*Note: When more than one factory, please report on page 2*Micromeritics Instrument Corporation  
4356 Communications Drive  
Norcross, GA, 30093  
United States of America Additional information on page 2

Ratings and principal characteristics

100/120/230Vac, 50/60Hz, 1250VA

Trademark (if any)



micromeritics

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

AutoChem III, Models 2930 and 2730

Additional information (if necessary may also be reported on page 2)

 Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010/AMD1:2016, IEC 61010-2-010:2019, IEC 61010-2-081:2019

**Comments:**

Refer to the Test Report for the Summary of Compliance with National Deviations

As shown in the Test Report Ref. No. which forms part of this Certificate

104803622ATL-001C, 104803622ATL-001D, 104803622ATL-001E

This CB Test Certificate is issued by the National Certification Body

Intertek Testing Services NA, Inc.  
545 E. Algonquin Road  
Arlington Heights  
IL 60005  
United States of America

Date: 2022-08-29

Signature: John Quigley 