



UK DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer:

Micromeritics Instrument Corporation
4356 Communications Drive
Norcross, GA 30093, USA

Hereby declares that the product:

MIC SAS II

is in conformity with the following UK legislation:

Electrical Equipment (Safety) Regulations 2016

Electromagnetic Compatibility Regulations 2016

Restriction of the Use of Certain Hazardous Substances in E&E Equipment Regulations 2012

and that the equipment is in conformity with the following designated and other appropriate standards;

Electrical Equipment (Safety) Regulations 2016

EN 61010-1:2010+A1:2019 - *Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements.*

EN IEC 61010-2-081:2020 – *Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes.*

Electromagnetic Compatibility Regulations 2016

EN 61326-1:2013 - *Electrical equipment for measurement, control and laboratory use — EMC requirements — Part 1: General requirements*

EN 61000-3-2:2014 - *Part 3-2: Limits — Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)*

EN 61000-3-3:2013 - *Part 3-3: Limits — Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection*

Restriction of the Use of Certain Hazardous Substances in E&E Equipment Regulations 2012

EN 63000:2018 - *Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances*

Name: Samuel Varner, Ph.D.

Title: Director, R&D

Signature: 

Date of issue: 2026-05-26

Location: Norcross, GA USA